

This work deals with testing of basic parameters of a contactless conductivity detector for capillary electrophoresis. The aim was to find area where the resulting signal dependence on concentration is linear, to determine the threshold sensitivity of the detector and verify that the detector is capable of providing reproducible results.